

НПЦ "ЭКО-СЕРВИС" САНКТ-ПЕТЕРБУРГ
TEST PARAMETERS FOR MINDRAY BS-120

TEST	FULL NAME	METHOD	DIRECTION	UNIT	DECIMAL	PRIM WAVE	SEC WAVE	SAMPLE uL	R1 VOL. uL	R2 VOL. uL	LINE LIMIT %	INCUBAT.	REACTION		SUBSTRATE EXHAUST	ANTIGEN SURPLUS	RESPONS		BLANK REAGT	R1 BLANK		MIX R. BLANK		LINEA		Standard Concn	FACTOR	SPEED
													START	END			LOWER	UPPER		LOWER	UPPER	LOWER	UPPER	LOWER	UPPER			
ALB	Albumin	Endpoint	Ascend	g/L	1	630	670	3	450	0	20	0	0	17	0	0	0	0	Yes	0	0	0	0	10	70	50		
ALP	Alcaline Phosphatase DGKC	Kinetic	Ascend	U/L	0	405		5	200	50	20	16	3	13	0	0	0	0	No	0	0	0	0	30	700	No	2757	
ALP	Alcaline Phosphatase IFCC	Kinetic	Ascend	U/L	0	405		5	200	50	20	16	3	13	0	0	0	0	No	0	0	0	0	30	700	No	2757	
ALT	ALAT-GPT	Kinetic	Descend	U/L	1	340		20	160	40	20	16	3	13	0	0	0	0	No	0	0	0	0	0	260	No	1750	
AMYL	Amylase	Kinetic	Ascend	U/L	0	405		5	200	50	20	16	3	12	0	0	0	0	No	0	0	0	0	28	1070	No	4554	
AST	ASAT-GOT	Kinetic	Descend	U/L	0	340		20	160	40	20	16	3	13	0	0	0	0	No	0	0	0	0	0	260	No	1750	
BILT	Bilirubin Total	Endpoint	Ascend	µmol/L	2	546	670	15	200	40	20	16	-1	17	0	0	0	0	Yes	0	0	0	0	0	513	71		
BILD	Bilirubin Direct	Endpoint	Ascend	µmol/L	2	546	700	20	160	40	20	16	-1	17	0	0	0	0	Yes	0	0	0	0	0	171	31		
CA	Calcium	Endpoint	Ascend	mmol/L	2	578	670	3	150	150	20	15	0	17	0	0	0	0	Yes	0	0	0	0	0	5	2,5		
CHOL	Cholesterol	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	0	0	33	0	0	0	0	Yes	0	0	0	0	0	19,4	5,17		
CL	Chloride	Endpoint	Ascend	mmol/L	1	510	670	3	300	0	20	0	0	20	0	0	0	0	Yes	0	0	0	0	80	130	100		
CREA	Creatinin	Fixed-Time	Ascend	µmol/L	0	510		30	150	150	20	16	2	8	0	0	0	0	Yes	0	0	0	0	30	1062	177		
GGT	Gamma Glutamyl	Kinetic	Ascend	U/L	0	405		20	160	40	20	16	3	13	0	0	0	0	No	0	0	0	0	8	230	No	1158	
GLU	Glucose GOD-POD	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	15	0	33	0	0	0	0	Yes	0	0	0	0	0	22	5,55		
GLU	Glucose HK	Endpoint	Ascend	mmol/L	2	340	670	3	300	0	20	15	0	33	0	0	0	0	Yes	0	0	0	0	0	40	5,55		
FE	Iron	Endpoint	Ascend	µmol/L	2	578	670	40	200	50	20	16	-1	17	0	0	0	0	Yes	0	0	0	0	3	179	17,9		
LDH	Lactate Dehydrogenase	Kinetic	Descend	U/L	0	340		3	200	50	20	16	3	13	0	0	0	0	No	0	0	0	0	150	1500	No	13400	
Mg	Magnesium	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	0	0	17	0	0	0	0	Yes	0	0	0	0	0	2,05	0,82		
PHOS	Inorganic Phosphorus	Endpoint	Ascend	mmol/L	2	340	670	3	300	0	20	15	0	33	0	0	0	0	Yes	0	0	0	0	0	4,8	1,61		
TP	Total Protein	Endpoint	Ascend	g/L	1	546	670	10	200	0	20	15	0	33	0	0	0	0	Yes	0	0	0	0	0	150	50		
TRIG	Triglycerides	Endpoint	Ascend	mmol/L	2	510	670	3	300	0	20	15	0	33	0	0	0	0	Yes	0	0	0	0	0	11,4	2,28		
UA	Uric Acid	Endpoint	Ascend	µmol/L	0	510	670	5	200	50	20	15	0	20	0	0	0	0	Yes	0	0	0	0	0	1428	357,00		
UREA Mono	Urea	Kinetic	Descend	mmol/L	2	340		3	300	0	20	15	2	7	0	0	0	0	No	0	0	0	0	0	33,3	8,33		